

Specifications of Automated External Defibrillator (AED)

S.N	Parameters	Technical Specifications	Remarks
1	Usage	To defibrillate Patient in Cardiac arrest	
2	Features	Biphasic	
		Easy Operation in 3 steps : Start-Analysis-shock	
		Automatic analysis on request	
		Manual override mode	
		Standard Energy Setting	
		LCD Display	
		Adult and Paediatric application	
3	Accessories	Rechargeable Battery	
		Memory Card	
		5 x Adult Pads and 1 x Pediatric Pads	
4	Terms and Conditions	Free installation and demo at Pokhara MH and Itahari MH	
		Warranty : Minimum 2 years/Provided by OEM whichever is greater	
		Manufacture Date must be less than 2 years at time of delivery.	
		Standard : ISO/ CE/ USFDA	

Handwritten signature and initials in green ink.

Handwritten signature and initials in green ink.

Handwritten signature and initials in green ink, with some illegible text below it.