

Specifications of Fogger Machine

S.N	Parameters	Technical Specifications	Remarks
1	Usage	To generate aerosol (Sub micron size) fog for disinfection of OT	
2	Power Supply	220V AC, 50 hz	
3	Spread & Reach	Minimum 40 ft. distance and 20 ft. height	
4	Space Treatment	Minimum 10000 cu. Ft.	
5	Disinfectant used	Maximum 400 ml for treatment of 1000 cu. Ft. area	
6	Nozzle Assembly	Non clogging design-Stainless Steel cap and barrel.	
7	Precision metering system	0-80ml/minute (Adjustable)	
8	Solutions tank & Lid	5 litre stainless steel grade 304 (autoclavable) with graduated markings Liquid Filling Nipple – SS 304 Tank Gasket – Silicon	
9	Silicon Tubing	Minimum 3 meters	
10	Accessories	Machine Handle, Mounting Ring, Strainer, Liquid Line Connectors, Hardware must be of Stainless Steel	
11	Intake air Filter	Two layer for dust & fog separation	
12	Terms and Conditions	The device must be brand new and date of manufacture must be less than 2 years on day of delivery. Installation and demonstration by suppliers at MH Pokhara and MH Itahari.	
13	Warranty	Minimum 1 year/ Provided by OEM whichever is greater	
14	Standard	ISO or Equivalent	







