

Specification of Scrub Station

S.N	Parameters	Technical Specifications	Remarks
1	Usage	Surgical scrub sink for use in OT complex.	
2	Material	16 gauge type 304/316 stainless steel high strength construction	
3	Power Supply	Power 220v 50 Hz	
4	Surface	Galvanized or mirror polished or chrome plated	
5	Backsplash	Minimum 4"-high (102mm)	
6	Shape	Squared/Rectangular front and sides.	
7	Front access panel	16 gauge type 304/316 stainless steel, removable	
8	Features	Temperature control mixing valve	
		Tap operated by foot IR	
		1.5" (38mm) Nominal Pipe Size stainless steel flat strainer drain with p-trap and tail piece.	
		Minimum 18" (457mm) flexible stainless steel supply hoses	
		Z-clips for mounting sink to wall or on a stand.	
		AC-DC control module gooseneck faucet, with waterproof connections with aerator.	
		Deck-mount faucet and Minimum 17.5" (444mm) front-to-back sink.	
		3-bay sinks, with or without partitions.	
9	Terms and Conditions	Free installation and demo at Pokhara MH and Itaheri MH is the responsibility of supplier	
		Warranty : Minimum 2 years/Provided by OEM whichever is greater	
		Manufacture Date must be less than 2 years at time of delivery.	
		Standard : ISO/CE Equivalent	

The bottom of the page contains several handwritten signatures and notes in green ink. On the left, there is a signature that appears to be 'S. K. B. S. S. S.' followed by some illegible text. In the center, there is a signature that looks like 'S. K. B. S. S. S.' and another one that is partially legible as 'S. K. B. S. S. S.'. On the right, there is a large signature that is mostly illegible, but it seems to be 'S. K. B. S. S. S.' followed by some other text. There are also some small marks and scribbles scattered around the signatures.