5	Features	Lead rubber apron, Waterproof	
6	Warranty	Minimum 1 year	
7	Terms and Conditions	The device must be brand new and date of manufacture must be less than 2 years on day of delivery.	
8	Standards	ISO / CE or equivalent	

Specification of Hanger 14" x 17", 14" x 14" and 10" x 12"

S.N	Parameters	Technical Specification	Remarks
1	Usage	To hang X-ray Films	
2	Material	Stainless Steel	
3	Warranty	Minimum 1 year	
4	Terms and Conditions	The device must be brand new and date of manufacture must be less than 2 years on day of delivery.	
5	Standards	ISO / CE or equivalent	

Specification of Lead bar for Xray Room

S.N	Parameters	Technical Specification	Remarks
1	Usage	To protect from X ray	
2	Movement	Wheeled and easily movable	
3	Thickness	Minimum 2 mm Lead protection apron	
4	Warranty	Minimum 1 year	
5	Terms and Conditions	The device must be brand new and date of manufacture must be less than 2 years on day of delivery.	
6	Standards	ISO / CE or equivalent	